

# Validation of the Omron HEM-7500F upper arm blood pressure monitor, in oscillometry mode, for clinic use and self measurement in a general population, according to the European Society of Hypertension International Protocol revision 2010

Hakuo Takahashi, Toyohiko Yokoi and Masamichi Yoshika

**Keywords:** blood pressure, device, European Society of Hypertension, guideline, measurement

*Department of Clinical Sciences and Laboratory Medicine, Kansai Medical University, Hirakata Osaka, Japan*

Correspondence to Professor Hakuo Takahashi, MD, PhD, Department of Clinical Sciences and Laboratory Medicine, Kansai Medical University, Shin-machi 2-3-1, Hirakata Osaka 573-1191, Japan

Tel: +81 72 804 2773; e-mail: takahash@hirakata.kmu.ac.jp

## Abstract

The Omron HEM-7500F, an upper arm blood pressure monitor, in oscillometry mode, for clinical use and self measurement, was validated, in a general population, according to the European Society of Hypertension International Protocol revision 2010. The protocol requirements were followed precisely. The device passed all of the requirements and, fulfilling the standards of the protocol, is recommended for clinical use.

## Device Details

Brand	Omron
Model	HEM-7500F
Manufacturer	Omron Healthcare Co.,Ltd
Location	Upper Arm
Method	Oscillometry
Purpose	Clinic Measurement, Self/ Home Measurement
Operation	Fully Automatic
Arm Cuff	Standard Adult: 17.0 cm to 36.0 cm
Other Features	The function to guide cuff wrapping, Memory capacity for 90 readings for two users,3 readings average value within 10 minutes,morning/evening average value, The indicator for blood pressure level, The graph function of blood pressure trend, The function to detect body motion, The function to detect irregular heartbeat, The function to display room temperature, Near field communication to connect with Personal computer and smartphone.



## Methodology

### Familiarisation

Hundreds of test-measurements were carried out. No problems were encountered.

### Recruitment

Hypertensive subjects were recruited from outpatients clinic in department of cardiology in the Knasai Medical University, Hirakata Hospital (Osaka, Japan). Some participated immediately without appointment. Normotensive subjects were recruited from outpatients and volunteers. There were some difficultlies in recruiting subjects with DBP in the high range.

**Screening and Recruitment Details**

Screening and Recruitment			Recruitment Ranges		
Total Screened	48		mmHg	All	On Rx
Total Excluded	15		< 90	1	0
Ranges Complete	0	Low	90 - 129	10	
Ranges Adjustment	0	SBP	Medium 130 - 160	11	6
Arrhythmias	6		High 161 - 180	9	1
Device Failure	0		> 180	2	
Poor Quality Sounds	1		< 40	0	1
Cuff Size Unavailable	1	Low	40 - 79	11	
Observer Disagreement	0	DBP	Medium 80 - 100	10	3
Distribution	0		High 101 - 130	12	
Other Reasons	7		> 130	0	3
Total Recruited	33				

**Procedure**

The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults was followed precisely.[1] Overseen by an independent supervisor, measurements were recorded by two observers blinded from both each other's readings and from the device readings.

**Results****Subject Details**

Sex		
Male : Female	21 : 12	
Age (years)		
Range (Low : High)	33 : 73	
Mean (SD)	49.8 (11.4)	
Arm Circumference (cm)		
Range (Low : High)	18.9 : 35.8	
Mean (SD)	28.5 (4.4)	
Cuff for test device		
Standard	33	(17.0 - 36.0 cm)
	SBP	DBP
Recruitment BP (mmHg)		
Range (Low : High)	85 : 188	49 : 126
Mean (SD)	142.6 (30.4)	88.9 (21.7)

**Observer Measurements in each Recruitment Range**

SBP (mmHg)	DBP (mmHg)
Overall Range (Low : High)	Overall Range (Low : High)
88 : 190	49 : 129
Low (< 130)	Low (< 80)
41	36
Medium (130 – 160)	Medium (80 – 100)
33	24
High (> 160)	High (> 100)
25	39
Maximum Difference	Maximum Difference
16	15


**Observer Differences**

	SBP (mmHg)	DBP (mmHg)	Repeated measurements
Observer 2 – Observer 1			
Range (Low : High)	-4 : +4	-4 : +2	
Mean (SD)	0.1 (1.4)	-0.4 (1.5)	0

**Validation Results**

Part 1	≤ 5 mmHg	≤ 10 mmHg	≤ 15 mmHg	Grade 1	Mean (mmHg)	SD (mmHg)
Pass Requirements						
Two of	73	87	96			
All of	65	81	93			
Achieved						
SBP	79	93	99	Pass	-1.0	4.5
DBP	78	98	99	Pass	-1.1	4.0
Part 2	2/3 ≤ 5 mmHg	0/3 ≤ 10 mmHg	Grade 2	Grade 3		
Pass Requirements						
	≥ 24	≤ 3				
Achieved						
SBP	28	0	Pass			Pass
DBP	28	2	Pass			Pass
Part 3						Result
						<b>PASS</b>

**Plots**



**Discussion**

Recruitment of subjects with high BP, particularly high DBP, proved to be difficult and accounted for most of the extra screened subjects; this is reflected in the overall distribution, as shown in the DBP plot, in which most of the points are below 115mmHg. The difference between observer and device was similar in the three BP ranges and all BP discrepancies were within 15mmHg.

**Conclusion**

As the device has reached the required standards, it is recommended for clinical and personal use in a general population.

**Acknowledgements and Conflict of Interest**

The monitor was supplied for the purposes of the study by the manufacturer OMRON Healthcare Co.,LTD. who also funded the study. None of the authors has any association with OMRON Healthcare Co.,LTD. or has received any personal benefit from OMRON Healthcare Co.,LTD.

## References

1. O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; on behalf of the Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the Validation of Blood Pressure Measuring Devices In Adults. *Blood Press Monit* 2010;15:23–38.