

## DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE [www.dableducational.org](http://www.dableducational.org) WEBSITE

### SECTION A - Please complete all items.

I **Mike Mai,** a Director of **Guangdong Transtek Medical Electronics Co.,Ltd**  
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

**Maker<sup>a</sup>** Guangdong Transtek Medical Electronics Co.,Ltd **Address** Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

**Manufacturer<sup>b</sup>** Artsana S.P.A **Address** Via Saldarini Catelli, 122070, Grandate(C)), Italy

**Brand<sup>c</sup>** Pic **Model<sup>d</sup>** easyRAPID

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

**Maker<sup>a</sup>** Guangdong Transtek Medical Electronics Co.,Ltd **Address** Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

**Manufacturer<sup>b</sup>** Guangdong Transtek Medical Electronics Co.,Ltd **Address** Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

**Brand<sup>c</sup>** TRANSTEK **Model<sup>d</sup>** TMB-1491

Existing validated blood pressure measuring device.

which has previously passed the ESH2010 protocol, the results of which were published as follows:

Hui Yong Tian, Si Jian Zeng, Xiao Yan Zhong, Wei Gong and Wen Jun Liu; Validation of TRANSTEK blood blood pressure monitor TMB-1491 for self-measurement according to the European Society of Hypertension International Protocol revision 2010, Blood Pressure Monitoring, 2015:280-282

Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1-18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A <sup>e</sup> <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>f</sup> <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A <sup>g</sup> <input checked="" type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- Provide the name and address of the actual maker of the device.
  - Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
  - Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
  - Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
  - Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
  - Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
  - Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

**SECTION B** An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

See attached document

**SECTION C** Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device\*
- An image of the screen layout of the device for which equivalence is being sought\*

\* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

**SECTION D** Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to [info@dableducational.org](mailto:info@dableducational.org).

Signature of Director Mike Mai

Company Stamp/Seal

Name Mike Mai

Date Oct. 29st, 2015

Signature of Witness Ada Zhang

Name Ada Zhang

Address Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437







**SECTION B of Declaration of Blood Pressure Measuring Device Equivalence**

	Existing Validated Device	Device applied for Validation
Model Name or Number	TMB-1491	easyRAPID
Casing		
Display		
Carrying/ Mounting Facilities	NO	
Software other than Algorithm	• Single User	• Single User
	• 60 sets memories	• 60 sets memories
	• WHO indicator	• WHO indicator
	• Low battery indicator	• Low battery indicator
	• Day/Time setting	• Day/Time setting

	<ul style="list-style-type: none"> <li>• Blood pressure &amp; heart rate measurement</li> </ul>	<ul style="list-style-type: none"> <li>• Blood pressure &amp; heart rate measurement</li> </ul>
	<ul style="list-style-type: none"> <li>• Kpa / mmHg unit</li> </ul>	<ul style="list-style-type: none"> <li>• mmHg unit</li> </ul>
	<ul style="list-style-type: none"> <li>• Blood pressure data memorized with date/time</li> </ul>	<ul style="list-style-type: none"> <li>• Blood pressure data memorized with date/time</li> </ul>
	<ul style="list-style-type: none"> <li>• Last 3 reading average</li> </ul>	<ul style="list-style-type: none"> <li>• Last 3 reading average</li> </ul>
	<ul style="list-style-type: none"> <li>• Error message indication</li> </ul>	<ul style="list-style-type: none"> <li>• Error message indication</li> </ul>
	<ul style="list-style-type: none"> <li>• Auto shut off when no operation for 1 min</li> </ul>	<ul style="list-style-type: none"> <li>• Auto shut off when no operation for 1 min</li> </ul>
Memory Capacity/ Number of stored measurements	60 sets(single user)	60 sets(single user)
Power Supply	4 x AAA	4x AAA or AC adaptor, output: 6VDC, 1A







Comparison of the PIC easyRAPID Automatic Blood Pressure Monitor with the Transtek TMB-1491

<p>Devices</p>	<p>PIC easyRAPID Automatic Blood Pressure Monitor</p>	<p>Transtek TMB-1491</p>
<p>Pictures</p>		
<p>Display</p>		
<p>Validation</p>		<p>ESH 2010</p>
<p>Device 1 Criteria</p>		<p><b>Measurement</b> Cuffs(Please state sizes and materials used) 22-32cm and 22-42cm</p>

		<p><b>Casing</b></p> <p><i>Appearance</i> color and shape different</p> <p><i>Ports</i> Cuff port</p> <p><i>Power</i> 4 x AAA battery only</p>
<p><b>Device 2 Criteria</b></p>	<p><b>Measurement</b> <i>Cuffs(Please state sizes and materials used)</i> 22-42cm</p> <p><b>Casing</b> <i>Appearance</i> color and shape different</p> <p><i>Ports</i> Cuff port and AC adaptor port</p> <p><i>Power</i> 4 x AAA battery, or AC adaptor, 6V 1A.</p>	
<p><b>Same Criteria</b></p>	<p><b>Measurement</b> <i>Accuracy</i> Pressure: 5°C-40°C within±0.4kPa(3mmHg) pulse value:±5%</p> <p><i>Method</i> : Oscillographic</p> <p><i>Ranges</i> Rated cuff pressure: 0mmHg~300mmHg Measurement pressure: 40mmHg-230mmHg pulse value: (40-199) beat/minute</p> <p><i>Inflation</i> Automatic Inflation Zero pressure check before inflation</p>	<p><b>Measurement</b> <i>Accuracy</i> Pressure: 5°C-40°C within±0.4kPa(3mmHg) pulse value:±5%</p> <p><i>Method</i>: Oscillographic</p> <p><i>Ranges</i> Rated cuff pressure: 0mmHg~300mmHg Measurement pressure: 40mmHg-230mmHg pulse value: (40-199) beat/minute</p> <p><i>Inflation</i> Automatic Inflation Zero pressure check before inflation</p>

	<p><i>Deflation</i>  <b>Automatic Deflation</b>                  Automatic safety release</p> <p><i>Cuffs (Please state sizes and materials used)</i>                  22-42cm, Polyester</p> <p><i>Sensors</i>                  Piezo-resistive</p> <p><i>Measurement Records</i>                  60</p> <p><i>Measurements other than Blood Pressure</i>                  Heart rate</p> <p><b>Buttons/Switches</b></p> <p><i>Power</i>                  Start/stop</p> <p><i>Function</i>                  SET button                  MEM button</p> <p><i>Analysis</i>                  N/A</p> <p><i>Event Marking</i>                  N/A</p> <p><i>Communication</i>                  N/A</p>	<p><i>Deflation</i>  <b>Automatic Deflation</b>                  Automatic safety release</p> <p><i>Cuffs(Please state sizes and materials used)</i>                  22-32cm and 22-42cm, Polyester</p> <p><i>Sensors</i>                  Piezo-resistive</p> <p><i>Measurement Records</i>                  60</p> <p><i>Measurements other than Blood Pressure</i>                  Heart rate</p> <p><b>Buttons/Switches</b></p> <p><i>Power</i>                  Start/stop</p> <p><i>Function</i>                  SET button                  MEM button</p> <p><i>Analysis</i>                  N/A</p> <p><i>Event Marking</i>                  N/A</p> <p><i>Communication</i></p>
--	---	---

<p><b>Display/Symbols/Indicators</b></p> <p><i>Preparation</i> Adjust to zero pressure</p>  <p><i>Measurement Procedure</i> Display the cuff pressure, heart rate symbol and measurement time</p>  <p><i>Post Measurement</i> Upper arm</p> <p><i>Date and Time</i> Display measurement time in the lower right corner of LCD</p> <p><i>Power</i> Low battery</p>	<p>N/A</p> <p><b>Display/Symbols/Indicators</b></p> <p><i>Preparation</i> Adjust to zero pressure</p>  <p><i>Measurement Procedure</i> Display the cuff pressure, heart rate symbol and measurement time</p>  <p><i>Post Measurement</i> Upper arm</p> <p><i>Date and Time</i> Display measurement time in the lower right corner of LCD</p> <p><i>Power</i> Low battery</p>
--	---



	<p><i>Function</i></p> <p>Measure blood pressure and heart rate Recall measurement records Delete measurement records</p> <p><i>Communication</i></p> <p>N/A</p> <p><i>Features</i></p> <p>Measuring during inflation</p> <p><b>Algorithms</b></p> <p><i>Averages and Differences</i></p> <p>Recall the average value of the last measurement</p> <p><i>Diagnostic</i></p> <p>N/A, indicate WHO blood pressure classification</p> <p><i>Functions</i></p> <p>Measure blood pressure and heart rate</p> <p><i>Communication</i></p> <p>N/A</p> <p><b>Casing</b></p> <p><i>Display</i></p> <p>LCD</p> <p><i>Ports</i></p> <p>Cuff port and AC adaptor port</p> <p><i>Power</i></p> <p>4*AAA battery, or AC adaptor, 6V 1A.</p> <p><i>Features</i></p> <p>ABS, trapezoid</p>	<p><i>Function</i></p> <p>Measure blood pressure and heart rate Recall measurement records Delete measurement records</p> <p><i>Communication</i></p> <p>N/A</p> <p><i>Features</i></p> <p>Measuring during inflation</p> <p><b>Algorithms</b></p> <p><i>Averages and Differences</i></p> <p>Recall the average value of the last measurement</p> <p><i>Diagnostic</i></p> <p>N/A, indicate WHO blood pressure classification</p> <p><i>Functions</i></p> <p>Measure blood pressure and heart rate</p> <p><i>Communication</i></p> <p>N/A</p> <p><b>Casing</b></p> <p><i>Display</i></p> <p>LCD</p> <p><i>Ports</i></p> <p>Cuff port</p> <p><i>Power</i></p> <p>4*AAA battery</p> <p><i>Features</i></p> <p>ABS, trapezoid</p>
--	---	--

<b>Comparable Criteria</b>	<p><i>Appearance</i> 110mm*110mm*40mm, color different</p> <p><i>Power</i> Except 4*AAA battery, also can be supplied by authorized AC adaptor</p> <p><i>Cuff size</i> 22-42cm</p>	<p><i>Appearance</i> 110mm*110mm*41mm, color different</p> <p><i>Power</i> Just supplied by 4*AAA battery</p> <p><i>Cuff size</i> 22-32cm and 22-42cm</p>
----------------------------	--	---

<b>Comments</b>	<i>Home Use Only, Self-measurement</i>
<b>Recommendation</b>	<b><i>Recommended</i></b>
<b>Date</b>	18 November 2015