

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2006

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items online.

I Takefumi Nakanishi Director of Omron Healthcare Europe B.V.
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

SEM-1 (HEM-7051-C12)
Blood pressure measuring device for which validation is claimed

blood pressure measuring device and the

Omron M3 Intellisense (HEM-7051-E)
Existing validated blood pressure measuring device

blood pressure measuring device, which has previously passed the International protocol, the results of which were published as follows

Asmar R, Khabouth J, Topouchian J, El Feghali R, Mattar J
Authors(s)

Validation of three automatic devices for self-measurement of blood pressure according to the International Protocol: The Omron M3 Intellisense (HEM-7051-E), the Omron M2

Compact (HEM 7102-E), and the Omron R3-I Plus (HEM 6022-E)
Title

Blood Pressure Monitoring 2010; 15:49-54
Publication Year Volume Pages

The only differences between the devices involve the following components:

(When a component is not relevant, both Yes and No should be left blank. Please provide details on any differences below.)

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	6	Cuff or Bladder	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>

Brief explanation of differences and further relevant details:

- 10) No clock setting button.
- 11) The symbol for irregular heartbeat detection is removed. Clock function is removed.
- 13) The function of irregular heartbeat detection is removed. Beeper is removed. Clock function is removed.
- 14) Stores 21 readings instead of 42.



SECTION B - Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original along with manuals for both devices to our address below.

Signature of Director T. Nakanishi

Name Takefumi Nakanishi

Date 08 Feb 2010

Signature of Witness J. Meijer

Name Janet Meijer

Address Omron Healthcare Europe B.V., Kruisweg 577, 2132NA Hoofddorp, The Netherlands

Company Stamp/Seal

OMRON HEALTHCARE EUROPE B.V.

Kruisweg 577




NL-2132 NA Hoofddorp

P.O. Box 2150 NL- 2130 GL Hoofddorp

Tel. +31 - 20 354 82 00

Fax +31 - 20 354 82 01

Comparison of the Omron SEM-1 (HEM-7051-C12) with the Omron M3 Intellisense (HEM-7051-E)

Devices	SEM-1 (HEM-7051-C12)	M3 Intellisense (HEM-7051-E)
Pictures		
Display		 <p> K. Systolic blood pressure L. Diastolic blood pressure M. Memory symbol N. Average value symbol O. Buzzer symbol P. Battery low symbol Q. Heartbeat symbol R. Pulse display S. Deflation symbol T. Date/Time display U. Irregular heartbeat symbol </p>
Validation		ESH
Device 1 Criteria		
Same Criteria	<p>Measurement</p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p> <p>Pulse accuracy ± 5% 1, 5</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>Pulse 40 bpm -180 bpm 1, 5</p> <p>Measurements are from single inflations 13</p> <p>Manually initiated measurements 13, 14</p> <p><i>Inflation</i></p> <p>Inflation 0 mmHg - 299 mmHg 1, 5, 7</p> <p>Automatic Inflation 7</p> <p>Fuzzy Logic^{Query 1} 7</p> <p>Press button if BP > 220 mmHg 7</p> <p>Manually adjustable inflation pressure 7</p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p>	<p>Measurement</p> <p><i>Accuracy</i></p> <p>BP accuracy ± 3 mmHg 1, 5</p> <p>Pulse accuracy ± 5% 1, 5</p> <p><i>Method</i></p> <p>Oscillometric measurement method 1, 5</p> <p>Pulse 40 bpm -180 bpm 1, 5</p> <p>Measurements are from single inflations 13</p> <p>Manually initiated measurements 13, 14</p> <p><i>Inflation</i></p> <p>Inflation 0 mmHg - 299 mmHg 1, 5, 7</p> <p>Automatic Inflation 7</p> <p>Fuzzy Logic 7</p> <p>Press button if BP > 220 mmHg 7</p> <p>Manually adjustable inflation pressure 7</p> <p><i>Deflation</i></p> <p>Automatic Deflation 8</p>

	Automatic safety release valve 8	Automatic safety release valve ^{Query 2} 8
	<i>Cuffs</i>	<i>Cuffs</i>
	Medium 146 mm × 446 mm (Arm circ. 22 to 32 cm) ^{Query 3} 6	Medium 146 mm × 446 mm (Arm circ. 22 to 32 cm) ^{Query 3} 6
	<i>Sensors</i>	<i>Sensors</i>
	Pressure sensor: capacitive 5	Pressure sensor: capacitive 5
	Buttons/Switches	Buttons/Switches
	<i>Measurement Records</i>	<i>Measurement Records</i>
	Memory 10	Memory 10
	Display/Symbols/Indicators	Display/Symbols/Indicators
	<i>Measurement Procedure</i>	<i>Measurement Procedure</i>
	Deflation symbol 11	Deflation symbol 11
	Heartbeat symbol during deflation 11	Heartbeat symbol during deflation 11
	<i>Post Measurement</i>	<i>Post Measurement</i>
	SBP, DBP and Pulse 11	SBP, DBP and Pulse 11
	Measurement error $EE, E, E/E$ and $E_{\alpha 25}$ 11	Measurement error $EE, E, E/E$ and $E_{\alpha 25}$ 11
	Hypertension (Blinking heartbeat) 11, 13	Hypertension (Blinking heartbeat) 11, 13
	Average symbol 11, 13	Average symbol 11, 13
	<i>Measurement Records</i>	<i>Measurement Records</i>
	Memory icon 11	Memory icon 11
	<i>Power</i>	<i>Power</i>
	Low battery 11, 17	Low battery 11, 17
	Algorithms	Algorithms
	<i>Diagnostic</i>	<i>Diagnostic</i>
	Normotension/Hypertension 13	Normotension/Hypertension 13
	135 / 85 mmHg thresholds 13	135 / 85 mmHg thresholds 13
	Case	Case
	<i>Display</i>	<i>Display</i>
	Single screen display 10	Single screen display 10
	Segment LCD 10	Segment LCD 10
	<i>Power</i>	<i>Power</i>
	4 “AA” batteries ~ 1500 measurements 17	4 “AA” batteries ~ 1500 measurements 17
	AC adapter (Optional) 17	AC adapter (Optional) 17
	Automatic switch-off when not used for 5 min 17	Automatic switch-off when not used for 5 min 17
Comparable Criteria	Measurement	Measurement
	<i>Measurement Records</i>	<i>Measurement Records</i>
	Memory: 21 measurements 14	Memory: 42 measurements 14
	Buttons/Switches	Buttons/Switches
	<i>Power</i>	<i>Power</i>
	On/Off with Start/Stop (Start/Stop Label) 10	On/Off with Start/Stop (O/I Start Label) 10

	<p>Display/Symbols/Indicators <i>Measurement Records</i> Memory recall number 11</p> <p>Algorithms <i>Averages</i> Last 3 measurements mean 13</p>	<p>Display/Symbols/Indicators <i>Date and Time</i> Date and Time (During memory recall) 11</p> <p>Algorithms <i>Averages</i> Last 3 measurements (within 10 min of each other) mean 13</p>
Device 2 Criteria		<p>Measurement <i>Cuffs</i> Large (Arm circ. 32-42 cm) (Optional) 6</p> <p>Buttons/Switches <i>Settings</i> Set 10</p> <p>Display/Symbols/Indicators <i>Measurement Procedure</i> Audible pulse indicator during deflation (Optional) 18 Beeps after measurement (Optional) 18</p> <p><i>Post Measurement</i> Irregular heartbeat 11, 13</p> <p><i>Date and Time</i> Date and Time 11</p> <p><i>Settings</i> Audible pulse indicator mode active 11, 18</p> <p>Algorithms <i>Diagnostic</i> Irregular heartbeat detection 13</p>
Web link		http://www.omron-healthcare.com/sitepreview.php?SiteID=227

Comments	<p>Query 1 Fuzzy logic: The manual, for the M3 Intellisense, states that fuzzy logic is used. It appears not to be available for the SEM1. There is no reference to this difference in the declaration. Please explain.</p> <p>Response 1 <i>The equivalent group of M3 Intellisense (HEM-7051-E) has the function of "Fuzzy logic", then SEM-1 also has Fuzzy logic as well in this case. However in our recent marketing approach some of models mention Fuzzy logic in the manual, some models do not mention, although all models in this group have Fuzzy logic. However we put the explanation of automatic inflation in each instruction manual for users to understand the function in spite of using the word of Fuzzy logic. As Fuzzy logic is related to Inflation mechanism, we checked "no differences" on the Part I - Item 7 of declaration forms.</i></p> <p>Query 2 Rapid pressure release: The manual, for the SEM1, include two deflation entries. In addition to the regular deflation, there is an automatic exhaust valve for rapid pressure release. This is understood to be a safety feature. It appears not to be available for</p>
-----------------	---

	<p>the M3 Intellisense. There is no reference to this difference in the declaration. Please explain.</p> <p>Response 2 <i>The fact we have is that the M3 Intellisense (HEM-7051-E) and the SEM1 (HEM-7051-C12) have same deflation mechanism. They have same valves for deflation system, as you mentioned, which are the regular deflation valve (slow deflation during measurement) and the rapid exhaust valve (release pressure rapidly from air system in the device after measurement to make comfortable and safe patients). Also these 2 valves are operated by automatic. In some device's manual e.g. M3 Intellisense (HEM-7051-E), we mention only "Deflation: Automatic pressure release valve" as one function of automatic deflation so that we could provide easy explanation to end users.</i></p> <p>Query 3 There appear to be some differences in the cuffs supplied with the monitors.</p> <p>There are different part numbers between those listed for the devices. No part numbers are provided for the SEM1 and no difference is made in the declaration. It is taken that there are no changes.</p> <p>Response 3 <i>These cuffs have no differences except cloth covers. The parts number difference comes from different cloth covers.</i></p>
Recommendation	The queries were adequately answered. Equivalence is recommended.
Date	26/08/2010