

# Validation of the Omron HEM-7252G-HP upper arm blood pressure monitor, in oscillometry mode, for clinic use and self measurement in a general population, according to the European Society of Hypertension International Protocol revision 2010

Hakuo Takahashi, Toyohiko Yokoi and Masamichi Yashika

**Keywords:** Blood pressure, European Society of Hypertension, guideline, device, measurement

*Department of Clinical Sciences and Laboratory Medicine, Kansai Medical University, Hirakata Osaka, Japan*

Correspondence to Hakuo Takahashi, MD, PhD, Department of Clinical Sciences and Laboratory Medicine, Kansai Medical University, Shin-machi 2-5-1, Hirakata Osaka 573-1010, Japan

Tel: +81 72 804 2691; e-mail: takahash@hirakata.kmu.ac.jp

## Abstract

The Omron HEM-7252G-HP, an upper arm blood pressure monitor, in oscillometry mode, for clinical use and self measurement, was validated, in a general population, according to the European Society of Hypertension International Protocol revision 2010. The protocol requirements were followed precisely. The device passed all of the requirements and, fulfilling the standards of the protocol, is recommended for clinical use.

## Device Details

Brand	Omron
Model	HEM-7252G-HP
Manufacturer	Omron Healthcare Co., Ltd.
Location	Upper Arm
Method	Oscillometry
Purpose	Clinic Measurement, Self/ Home Measurement
Operation	Fully Automatic
Arm Cuffs	Small Adult: 17.0 cm to 22.0 cm, Standard Adult: 22.0 cm to 32.0 cm and Large Adult: 32.0 cm to 42.0 cm
Other Features	The function to guide cuff wrapping, to detect body movement, to detect irregular heart beat, and to measure room temperature Memory capacity for 255 readings Third generation of mobile telecommunications technology to connect with the dedicated server The function to measure blood pressure during user sleep



---

## Methodology

### Familiarisation

Hundreds of test-measurements were carried out. No problems were encountered.

### Recruitment

Hypertensive subjects were recruited from outpatients clinic in the Department of Cardiology in Kansai Medical University, Hirakata Hospital (Osaka, Japan). Some participated immediately without appointment. Normotensive subjects were recruited from outpatients and volunteers.

**Screening and Recruitment Details**

Screening and Recruitment			Recruitment Ranges		
Total Screened	44		mmHg	All	On Rx
Total Excluded	11		< 90	0	0
Ranges Complete	0		Low	90 - 129	12
Ranges Adjustment	0	SBP	Medium	130 - 160	11
Arrhythmias	4		High	161 - 180	6
Device Failure	0		> 180	4	2
Poor Quality Sounds	1				
Cuff Size Unavailable	0		< 40	0	0
Observer Disagreement	0		Low	40 - 79	11
Distribution	0	DBP	Medium	80 - 100	11
Other Reasons	6		High	101 - 130	11
Total Recruited	33		> 130	0	2

**Procedure**

The European Society of Hypertension International Protocol revision 2010 for the validation of blood pressure measuring devices in adults was followed precisely.[1] Overseen by an independent supervisor, measurements were recorded by two observers blinded from both each other's readings and from the device readings.

**Results****Subject Details**

Sex		
Male : Female	19 : 14	
Age (years)		
Range (Low : High)	28 : 78	
Mean (SD)	49.7 (11.6)	
Arm Circumference (cm)		
Range (Low : High)	20.1 : 37.2	
Mean (SD)	28.1 (4.6)	
Cuff for test device		
Small	4	(17.0 - 22.0 cm)
Standard	23	(22.0 - 32.0 cm)
Large	6	(32.0 - 42.0 cm)
	SBP	DBP
Recruitment BP (mmHg)		
Range (Low : High)	90 : 207	48 : 129
Mean (SD)	144.2 (32.7)	89.8 (20.2)

**Observer Measurements in each Recruitment Range**

SBP (mmHg)		DBP (mmHg)	
Overall Range (Low : High)	86 : 218	Overall Range (Low : High)	48 : 134
Low (< 130)	41	Low (< 80)	26
Medium (130 – 160)	28	Medium (80 – 100)	42
High (> 160)	30	High (> 100)	31
Maximum Difference	13	Maximum Difference	16


**Observer Differences**

	SBP (mmHg)	DBP (mmHg)	Repeated measurements
Observer 2 – Observer 1			
Range (Low : High)	-2 : +4	-2 : +4	
Mean (SD)	-0.1 (1.2)	0.1 (1.3)	0

**Validation Results**

Part 1	≤ 5 mmHg	≤ 10 mmHg	≤ 15 mmHg	Grade 1	Mean (mmHg)	SD (mmHg)
Pass Requirements						
Two of	73	87	96			
All of	65	81	93			
Achieved						
SBP	76	92	97	Pass	-1.5	5.1
DBP	83	97	99	Pass	-1.2	3.9
Part 2	2/3 ≤ 5 mmHg	0/3 ≤ 10 mmHg	Grade 2	Grade 3		
Pass Requirements						
	≥ 24	≤ 3				
Achieved						
SBP	28	1	Pass			Pass
DBP	29	2	Pass			Pass
Part 3						Result
						<b>PASS</b>

**Plots**



**Discussion**

No specific problems were encountered during validation and distribution conditions were fulfilled. But it was difficult to fulfill requirements of the International Protocol because some subjects were very hypertensive over SBP 200mmHg. The agreement between observer and device was similar in the three BP ranges and almost BP discrepancies were within 15mmHg.

**Conclusion**

As the device has reached the required standards, it is recommended for clinical and personal use in a general population.

**Acknowledgements and Conflict of Interest**

The monitor was supplied for the purposes of the study by the manufacturer OMRON Healthcare CO., LTD. who also funded the study. None of the authors has any association with OMRON Healthcare CO., LTD. or has received any personal benefit from OMRON Healthcare CO., LTD.

## References

1. O'Brien E, Atkins N, Stergiou G, Karpettas N, Parati G, Asmar R, Imai Y, Wang J, Mengden T, Shennan A; on behalf of the Working Group on Blood Pressure Monitoring of the European Society of Hypertension. European Society of Hypertension International Protocol revision 2010 for the Validation of Blood Pressure Measuring Devices In Adults. *Blood Press Monit* 2010;15:23–38.