

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Liu Zhiqing**, a Director of **Andon Health Co.,Ltd.**,
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a Peroxfarma, S.A. **Address** Carrer Provença 328. 08037 Barcelona (Spain)
Manufacturer^b Andon **Address** No.3 JinPing Street, YaAn Road, Nankai District, Tianjin 300190, China.
Brand^c ICO **Model^d** MT-20

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a Andon **Address** No.3 JinPing Street, YaAn Road, Nankai District, Tianjin 300190, China.
Manufacturer^b Andon **Address** No.3 JinPing Street, YaAn Road, Nankai District, Tianjin 300190, China.
Brand^c Andon **Model^d** KD-5965

Existing validated blood pressure measuring device.

which has previously passed the ESH 2010 protocol, the results of which were published as follows:

Validation of the Andon KD-5965 upper-arm blood pressure monitor for home blood pressure monitoring according to the European Society of Hypertension International Protocol revision 2010
Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1-18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^g <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
 - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
 - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
 - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
 - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
 - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
 - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

- (9) The Model Name is changed to ICO MT-20 from Andon KD-5965;
- (10) The new device has a different industrial design.
- (11) please see the attachment.
- (13) Just changing the way of the code written, not the software scheme
- (14) Stores 2*60 readings instead of 60 readings.

SECTION C Please check that the following are included with the application

- A manual for the validated device
 - A manual for the device for which equivalence is being sought
 - Completed DET9 Form
 - An image of the device for which equivalence is being sought
 - An image of the screen layout of validated device*
 - An image of the screen layout of the device for which equivalence is being sought*
- * Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director 

Name Liu Zhiqing

Date May 12, 2022





Signature of Witness 

Name SunGuimei

Address No.3 JinPing Street, YaAn Road, Nankai District, Tianjin 300190, China



Comparison of the ICO MT-20 with the Andon KD-5965

Devices – Item 9	ICO MT-20	Andon KD-5965
Pictures		
Display Image		
Validation	N/A	ESH 2010
Category	SBPM	SBPM
Casing – Item 10	<p><i>Dimensions</i></p> <p>118*97*175cm</p> <p><i>Ports</i></p> <p>Cuff port and Adapter port</p>	<p><i>Dimensions</i></p> <p>177*100*118</p> <p><i>Ports</i></p> <p>Cuff port and Adapter port</p>

	<i>Features</i> N/A	<i>Features</i> N/A
Display – Item 11	<i>Type</i> Segment LCD	<i>Type</i> Segment LCD
Carrying/Mounting Facilities – Item 12	N/A	N/A
Software other than Algorithm – Item 13	Just changing the way of the code written, not the software scheme	N/A
Memory Capacity Item 14	2*60	60
Printing Facilities Item 15	N/A	N/A
Communication Facilities – Item 16	N/A	N/A
Power Supply Item 17	4*size AA	4*size AA
Other differences	<i>Other Details on Equivalent device that are different to Validated device</i> N/A	<i>Other Details on Validated device that are different to Equivalent device</i> N/A
Same Criteria	<i>Measurement</i> <i>Accuracy</i> Pressure: $\pm 3\text{mmHg}$ Pulse rate: Less than 60: $\pm 3\text{bpm}$ More than 60 (incl.) : $\pm 5\%$	<i>Measurement</i> <i>Accuracy</i> Pressure: $\pm 3\text{mmHg}$ Pulse rate: Less than 60: $\pm 3\text{bpm}$ More than 60 (incl.) : $\pm 5\%$

	<p><i>Method</i></p> <p>Oscillometric</p> <p><i>Ranges</i></p> <p>Cuff pressure: 0-300mmHg</p> <p>Systolic: 60-260mmHg</p> <p>Diastolic: 40-199mmHg</p> <p>Pulse rate: 40-180 beats/minute</p> <p><i>Inflation</i></p> <p>Automatic inflation by internal pump</p> <p><i>Deflation</i></p> <p>Automatic speed deflation system</p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p>22-30cm (identical to 20-34cm,only silk mark is different)</p> <p>30-42cm (identical to 30-44cm,only silk mark is different)</p> <p>42-48cm (identical to 40-48cm,only silk mark is different)</p> <p>Materials are Nylon and polyester</p> <p><i>Sensors</i></p> <p>KD-2107-006GA</p> <p><i>Measurement Records</i></p> <p>2*60 readings</p>	<p><i>Method</i></p> <p>Oscillometric</p> <p><i>Ranges</i></p> <p>Cuff pressure: 0-300mmHg</p> <p>Systolic: 60-260mmHg</p> <p>Diastolic: 40-199mmHg</p> <p>Pulse rate: 40-180 beats/minute</p> <p><i>Inflation</i></p> <p>Automatic inflation by internal pump</p> <p><i>Deflation</i></p> <p>Automatic speed deflation system</p> <p><i>Cuffs(Please state sizes and materials used)</i></p> <p>15-24cm</p> <p>20-34cm</p> <p>30-44cm</p> <p>40-48cm</p> <p>Materials are Nylon and polyester</p> <p><i>Sensors</i></p> <p>KD-2107-006GA</p> <p><i>Measurement Records</i></p> <p>60 readings</p>
--	---	--

	<p><i>Measurements other than Blood Pressure</i></p> <p>Pulse rate and IHB</p> <p>Buttons/Switches</p> <p><i>Power</i></p> <p>Start/Stop button</p> <p><i>Measurement Records</i></p> <p>Memory button M1,M2</p> <p><i>Function</i></p> <p>Date and Time setting</p> <p><i>Analysis</i></p> <p>N/A</p> <p><i>Event Marking</i></p> <p>N/A</p> <p><i>Communication</i></p> <p>N/A</p> <p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p>N/A</p> <p><i>Measurement Procedure</i></p> <p>Measuring during deflation</p>	<p><i>Measurements other than Blood Pressure</i></p> <p>Pulse rate and IHB</p> <p>Buttons/Switches</p> <p><i>Power</i></p> <p>Start/Stop button</p> <p><i>Measurement Records</i></p> <p>Memory button MEM</p> <p><i>Function</i></p> <p>Date and Time setting</p> <p><i>Analysis</i></p> <p>N/A</p> <p><i>Event Marking</i></p> <p>N/A</p> <p><i>Communication</i></p> <p>N/A</p> <p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p>N/A</p> <p><i>Measurement Procedure</i></p> <p>Measuring during deflation</p>
--	---	---

	<p><i>Post Measurement</i></p> <p>Upper arm</p> <p><i>Measurement Records</i></p> <p>2*60 readings</p> <p><i>Date and Time</i></p> <p>Displayed on LCD</p> <p><i>Power</i></p> <p>4*size AA</p> <p><i>Function</i></p> <p>N/A</p> <p><i>Communication</i></p> <p>N/A</p> <p><i>Features</i></p> <p>N/A</p> <p><i>Not described</i></p> <p>N/A</p> <p>Algorithms</p> <p><i>Averages and Differences</i></p> <p>No such function</p>	<p><i>Post Measurement</i></p> <p>Upper arm</p> <p><i>Measurement Records</i></p> <p>60 readings</p> <p><i>Date and Time</i></p> <p>Displayed on LCD</p> <p><i>Power</i></p> <p>4*size AA</p> <p><i>Function</i></p> <p>N/A</p> <p><i>Communication</i></p> <p>N/A</p> <p><i>Features</i></p> <p>N/A</p> <p><i>Not described</i></p> <p>N/A</p> <p>Algorithms</p> <p><i>Averages and Differences</i></p> <p>No such function</p>
--	---	---

	<p><i>Diagnostic</i></p> <p>N/A</p> <p><i>Functions</i></p> <p>N/A</p> <p><i>Communication</i></p> <p>N/A</p>	<p><i>Diagnostic</i></p> <p>N/A</p> <p><i>Functions</i></p> <p>N/A</p> <p><i>Communication</i></p> <p>N/A</p>
Comparable Criteria		

Comments		
Recommendation	Recommended	
Date	May 2022	