

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I **Kevin Tan**, a Director of **Guangdong Transtek Medical Electronics Co.,Ltd** ,
Name of a Company Director Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Manufacturer^b **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Brand^c **Alvita/Kinetik Wellbeing** Model^d **TMB-1117-S**
 Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Manufacturer^b **Guangdong Transtek Medical Electronics Co.,Ltd** Address **Zone A, No.105 ,Dongli Road, Torch Development District, Zhongshan,528437,Guangdong,China**

Brand^c **TRANSTEK** Model^d **TMB-988**
 Existing validated blood pressure measuring device.

which has previously passed the **ESH2002** protocol, the results of which were published as follows:

Title: **Validation of the Transtek TMB-988 wrist blood pressure monitor for home blood pressure monitoring according to the International Protocol.**

Author: **Tian HY, Liu WJ, Li SG, Song Z, Gong W.**

Publication: **Blood Press Monit 2010;15(6):326-8 doi: 10.1097/MBP.0b013e32833f56fb**
 Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1–18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>

An explanation of each item ticked “Yes” must be included in Section B or on a separate sheet.

Notes: a Provide the name and address of the actual maker of the device.
 b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.

- c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
- d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
- e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
- f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
- g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

See attached document

SECTION C Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device*
- An image of the screen layout of the device for which equivalence is being sought*

* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director Kevin Tan

Company Stamp/Seal

Name Kevin Tan

Date 8 April, 2022





Signature of Witness Caroline Liu



Name Caroline.liu

Address Zone A, No.105, Dongli Road, Torch Development District,
528437 Zhongshan, Guangdong, China

Comparison of the Alvita/Kinetik Wellbeing TMB-1117-S with the Guangdong Transtek Medical TMB-988

Devices – Item 9	Alvita/Kinetik Wellbeing TMB-1117-S	Guangdong Transtek Medical TMB-988
Pictures		
Display Image		
Validation		ESH 2010
Category	<i>wrist device for self measurement of blood pressure</i>	<i>wrist device for self measurement of blood pressure</i>
Casing – Item 10	<i>Dimensions</i> <i>62mm*75mm*31mm</i>	<i>Dimensions</i> <i>73mm*67.5mm*22.5mm</i>

© 2002-2022 dabl® Educational Trust Limited – No reproduction of this document is permitted without the written authorisation of dabl® Educational Trust Limited dabl® Educational Trust Limited is a not-for-profit organisation.

	<p><i>Ports</i></p> <p><i>Cuff port</i></p> <p><i>Features</i></p> <p>Alvita printing</p> <p>Blood pressure measurement</p> <p>WHO classification</p> <p>Pulse rate</p> <p>Button printing</p>	<p><i>Ports</i></p> <p><i>Cuff port</i></p> <p><i>Features</i></p> <p>Transtek printing</p> <p>Blood pressure measurement</p> <p>WHO classification</p> <p>Pulse rate</p> <p>Button printing</p>
Display – Item 11	<p><i>Type</i></p> <p>LCD</p> <p>LCD V.A32mm*44mm</p>	<p><i>Type</i></p> <p>LCD</p> <p>LCD V.A.35mm×41mm</p>
Carrying/Mounting Facilities – Item 12	None	None
Software other than Algorithm – Item 13	<p>One User</p> <p>60 sets memories/per user</p> <p>2 grade indicator</p> <p>mmHg unit</p>	<p>Dual Users</p> <p>60 sets memories/per user</p> <p>2 grade indicator</p> <p>mmHg unit</p>
Memory Capacity Item 14	60 sets memories/one user	60 sets memories/two user
Printing Facilities Item 15	N/A	N/A
Communication Facilities – Item 16	N/A	N/A

<p>Power Supply Item 17</p>	<p>2 dry cells 1.5V AAA</p>	<p>2 dry cells 1.5V AAA</p>
<p>Other differences</p>	<p><i>Other Details on Equivalent device that are different to Validated device</i> N/A</p>	<p><i>Other Details on Validated device that are different to Equivalent device</i> N/A</p>
<p>Same Criteria</p>	<p>Measurement</p> <p><i>Accuracy</i></p> <p>5°C-40°C within±3mmHg</p> <p>Pulse value:±5%</p> <p><i>Method</i></p> <p><i>Oscillographic testing mode</i></p> <p><i>Ranges</i></p> <p><i>Rated cuff pressure:</i></p> <p><i>Pressure: 0mmHg~299mmHg</i></p> <p><i>Pulse value: (40-199)beat/minute</i></p> <p><i>Inflation</i></p> <p><i>Automatic inflation</i></p> <p><i>Deflation</i></p> <p><i>Automatic deflation</i></p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p>13.5CM-21.5CM,polyester</p>	<p>Measurement</p> <p><i>Accuracy</i></p> <p>5°C-40°C within±3mmHg(0.4kPa)</p> <p>Pulse value:±5%</p> <p><i>Method</i></p> <p><i>Oscillographic testing mode</i></p> <p><i>Ranges</i></p> <p><i>Rated cuff pressure:</i></p> <p><i>Pressure: 0kpa – 39.9kpa (0mmHg~299mmHg)</i></p> <p><i>Pulse value: (40-199) beat/minute</i></p> <p><i>Inflation</i></p> <p><i>Automatic inflation</i></p> <p><i>Deflation</i></p> <p><i>Automatic deflation</i></p> <p><i>Cuffs(Please state sizes and materials used)</i></p> <p>13.5CM-21.5CM polyester</p>

	<p><i>Sensors</i></p> <p><i>Piezo-resistive</i></p> <p><i>Measurements other than Blood Pressure</i></p> <p><i>Pulse rate</i></p> <p>Buttons/Switches</p> <p><i>Power button – ‘ON SYMBOL’ button</i></p> <p><i>Memory button – MEM button</i></p> <p><i>Set button – SET button</i></p> <p><i>Analysis – N/A</i></p> <p><i>Event Marking – N/A</i></p> <p><i>Communication – N/A</i></p> <p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p><i>Automatic Zero setting</i></p> <p><i>Measurement Procedure</i></p> <p><i>Inflation symbol</i></p> <p><i>Pressure value indication</i></p> <p><i>Current time</i></p> <p><i>Measurement Records</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p> <p><i>Memory Query symbol</i></p> <p><i>Power</i></p>	<p><i>Sensors</i></p> <p><i>Piezo-resistive</i></p> <p><i>Measurements other than Blood Pressure</i></p> <p><i>Pulse rate</i></p> <p>Buttons/Switches</p> <p><i>Power button – START/STOP button</i></p> <p><i>Memory button – MEM button</i></p> <p><i>Set button – SET button</i></p> <p><i>Analysis – N/A</i></p> <p><i>Event Marking – N/A</i></p> <p><i>Communication – N/A</i></p> <p>Display/Symbols/Indicators</p> <p><i>Preparation</i></p> <p><i>Automatic Zero setting</i></p> <p><i>Measurement Procedure</i></p> <p><i>Inflation symbol</i></p> <p><i>Pressure value indication</i></p> <p><i>Current time</i></p> <p><i>Measurement Records</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p> <p><i>Measurement time</i></p> <p><i>Memory Query symbol</i></p> <p><i>Power</i></p>
--	---	--

	<p><i>Low power</i></p> <p><i>Features</i></p> <p><i>Measuring during inflation</i></p> <p>Algorithms</p> <p><i>Equivalent device has the identical measurement algorithm as the validated device.</i></p>	<p><i>Low power</i></p> <p><i>Features</i></p> <p><i>Measuring during inflation</i></p> <p>Algorithms</p> <p><i>Equivalent device has the identical measurement algorithm as the validated device.</i></p>
Comparable Criteria	<p>Measurement</p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p><i>About 13.5cm-21.5cm, polyester</i></p> <p><i>Measurement Records</i></p> <p><i>60 sets/per user, total one user</i></p> <p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p>	<p>Measurement</p> <p><i>Cuffs (Please state sizes and materials used)</i></p> <p><i>About 13.5cm-21.5cm, polyester</i></p> <p><i>Measurement Records</i></p> <p><i>60 sets/per user, total two users</i></p> <p>Display/Symbols/Indicators</p> <p><i>Post Measurement</i></p> <p><i>Systolic blood pressure (SYS)</i></p> <p><i>Diastolic blood pressure (DIA)</i></p> <p><i>Pulse rate</i></p>
Comments		
Recommendation	RECOMMENDED	
Date	May 2022	