

DECLARATION OF BLOOD PRESSURE MEASURING DEVICE EQUIVALENCE 2013

A SIGNED COPY WILL BE POSTED ON THE www.dableducational.org WEBSITE

SECTION A - Please complete all items.

I Mike Mai, a Director of Guangdong Transtek Medical Electronics Co., Ltd, Company name

hereby state that there are no differences that will affect blood pressure measuring accuracy between the

Maker^a Guangdong Transtek Medical Electronics Co., Ltd Address Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

Manufacturer^b PIKDARE S.r.l. Address Via Saldarini Catelli, 10 - 22070, Casnate con Bernate (CO), Italy

Brand^c PIC Solution Model^d mobileRAPID

Blood pressure measuring device for which validation is claimed. If alternative model names are used, include all.

blood pressure measuring device and the validated blood pressure measuring device

Maker^a Guangdong Transtek Medical Electronics Co., Ltd Address Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

Manufacturer^b Guangdong Transtek Medical Electronics Co., Ltd Address Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

Brand^c TRANSTEK Model^d LS808-B

Existing validated blood pressure measuring device.

which has previously passed the ESH2010 protocol, the results of which were published as follows:

Zhong HuaLiu, Xian Yue Liu, Wen Jun Wu Validation of Transtek LS808-B for self/home measurement according to the European Society of Hypertension International Protocol reversion 2010 Full reference

The only differences between the devices involve the following components:

Tick one box for each item 1-18.

Part I	1	Algorithm for Oscillometric Measurements	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^e <input type="checkbox"/>
	2	Algorithm for Auscultatory Measurements	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	3	Artefact/Error Detection	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	4	Microphone(s)	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^f <input checked="" type="checkbox"/>
	5	Pressure Transducer	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	6	Cuffs or Bladders	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	7	Inflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
	8	Deflation Mechanism	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	
Part II	9	Model Name or Number	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	10	Casing	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	11	Display	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	12	Carrying/Mounting Facilities	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	13	Software other than Algorithm	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	14	Memory Capacity/Number of stored measurements	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	15	Printing Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	16	Communication Facilities	Yes <input type="checkbox"/>	No <input type="checkbox"/>	N/A ^g <input checked="" type="checkbox"/>
	17	Power Supply	Yes <input checked="" type="checkbox"/>	No <input type="checkbox"/>	
	18	Other Facilities	Yes <input type="checkbox"/>	No <input checked="" type="checkbox"/>	N/A ^g <input type="checkbox"/>

An explanation of each item ticked "Yes" must be included in Section B or on a separate sheet.

- Notes:
- a Provide the name and address of the actual maker of the device.
 - b Provide the name and address of the legal manufacturer of the device, even if it is the same as that of the maker.
 - c Provide the name of the brand under which it is sold, even if it is the same as that of the manufacturer or maker.
 - d Provide the model name. If alternative or internal model names are used, include all. Each device must be uniquely identifiable.
 - e Only tick N/A (Not Applicable) if neither device measures blood pressure using the oscillometric method.
 - f Only tick N/A (Not Applicable) if neither device measures blood pressure using the auscultatory method.
 - g Only tick N/A (Not Applicable) if neither device provides printing, communication or other facilities, as appropriate.

SECTION B An explanation for each item, 1 to 18, ticked "Yes" in Section A must be provided here or in an attached document. All differences between the devices must be described.

See attached document

SECTION C Please check that the following are included with the application

- A manual for the validated device
- A manual for the device for which equivalence is being sought
- An image of the validated device
- An image of the device for which equivalence is being sought
- An image of the screen layout of validated device*
- An image of the screen layout of the device for which equivalence is being sought*

* Screen layouts shown complete, and without obscuring labels or lines, in manuals need not be included separately.

SECTION D Complete all items, bar signatures and seal, online and print. Sign and seal it then send the original to our address below. Please email a signed copy of this form, together with the manuals and images for both devices, to info@dableducational.org.

Signature of Director _____

Company Stamp/Seal

Name Mike Mai





Date 2017.05.27 _____

Signature of Witness 譚文昊 _____

Name





Address Zone A, No.105, Dongli Rd., Torch Development District, Zhongshan, Guangdong, China, 528437

Comparison of the PIKDARE PiC Solution mobileRAPID with the Transtek LS808-B

Devices – Item 9	PIKDARE PiC Solution mobileRAPID	Transtek LS808-B
Pictures		
Display Image		
Validation		ESH 2010 ESH 2002 BHS AAMI


Category		
Casing – Item 10	<p><i>Dimensions</i> 133.0mm*74.3mm*25.4mm</p> <p><i>Ports</i> Cuff port and DC power port</p> <p><i>Features</i> 133.0mm*74.3mm*25.4mm 22-42cm</p>	<p><i>Dimensions</i> 130.9mm*73mm*29.4mm</p> <p><i>Ports</i> Cuff port and DC power port</p> <p><i>Features</i> 130.9mm*73mm*29.4mm 22-42cm</p>
Display – Item 11	<p><i>Indication</i> Unit: mmHg unit ESH indicator</p> <p><i>Measurement records:</i> Maximum 100 records per each user(dual users) If the measurement results beyond the measuring range.</p>	<p><i>Indication</i> Unit: Kpa / mmHg unit WHO indicator</p> <p><i>Measurement records:</i> Maximum 60 records per each user(dual users) If the measurement results beyond the measuring range ,display shows OUT.</p>
Carrying/Mounting Facilities – Item 12	Blue-black bag	NA
Software other than Algorithm – Item 13	<p><i>Dual Users</i> 200 sets memories ESH indicator mmHg unit</p>	<p><i>Dual Users</i> 120sets memories WHO indicator</p>
Memory Capacity Item 14	200 sets memories	120 sets memories
Printing Facilities Item 15	NA	NA
Communication Facilities – Item 16	NA	NA
Power Supply Item 17	<ol style="list-style-type: none"> lithium battery 6V DC Jack 	<ol style="list-style-type: none"> lithium battery 6V DC Jack
Other differences	<p><i>Other Details on Equivalent device that are different to Validated device</i> N/A</p>	<p><i>Other Details on Validated device that are different to Equivalent device</i> N/A</p>
Same Criteria	Measurement Accuracy	Measurement Accuracy

	<p>Pressure: ± 3 mmHg Pulse value: $\pm 5\%$</p> <p><i>Method</i> Oscillography</p> <p><i>Ranges</i> Rated cuff pressure:0mmHg~300mmHg Measurement pressure:40mmHg~230mmHg Pulse value:(40-199) beat/minute</p> <p><i>Inflation</i> Automatic Deflation Zero pressure check before inflation</p> <p><i>Deflation</i> Automatic Deflation Automatic safety release</p> <p><i>Cuffs (Please state sizes and materials used)</i> 22-42cm, Polyester</p> <p><i>Sensors</i> Piezo-resistive</p> <p><i>Measurement Records</i> Last 3 reading average</p> <p>Buttons/Switches <i>Power</i> User 1 / User 2/ MEM</p> <p><i>Function</i> User 1 / User 2/ MEM/"lock" button</p> <p>Casing</p> <p><i>Ports</i> Cuff port and DC power port</p> <p><i>Power</i> 3.7V lithium battery</p>	<p>Pressure: ± 3 mmHg Pulse value: $\pm 5\%$</p> <p><i>Method</i> Oscillography</p> <p><i>Ranges</i> Rated cuff pressure:0mmHg~300mmHg Measurement pressure:40mmHg~230mmHg Pulse value:(40-199) beat/minute</p> <p><i>Inflation</i> Automatic Deflation Zero pressure check before inflation</p> <p><i>Deflation</i> Automatic Deflation Automatic safety release</p> <p><i>Cuffs (Please state sizes and materials used)</i> 22-42cm, Polyester</p> <p><i>Sensors</i> Piezo-resistive</p> <p><i>Measurement Records</i> Last 3 reading average</p> <p>Buttons/Switches <i>Power</i> User 1 / User 2/ MEM</p> <p><i>Function</i> User 1 / User 2/ MEM/"lock" button</p> <p>Casing</p> <p><i>Ports</i> Cuff port and DC power port</p> <p><i>Power</i> 3.7V lithium battery</p>
--	---	---


	<p>Display/Symbols/Indicators <i>Preparation</i></p>  <p><i>Measurement Procedure</i></p>  <p><i>Measurement Records</i></p>	<p>Display/Symbols/Indicators <i>Preparation</i></p>  <p><i>Measurement Procedure</i></p>  <p><i>Measurement Records</i></p>
--	--	--



Date and Time




Date and Time



Power
Low power

Function
Measure blood pressure and heart rate
Recall measurement records
Delete measurement records

Communication



Power

Low power
Function
Measure blood pressure and heart rate
Recall measurement records
Delete measurement records

Communication



Features
Measuring during inflation

Not described

Algorithms

Averages and Differences
Averages and Differences
Recall the average value of last measurement

Diagnostic
N/A

Functions
Measure blood pressure and heart rate



Features
Measuring during inflation

Not described

Algorithms

Averages and Differences
Averages and Differences
Recall the average value of last measurement

Diagnostic
N/A

Functions
Measure blood pressure and heart rate

	<i>Communication</i>	<i>Communication</i>
Comparable Criteria		

Comments		
Recommendation	Recommended	
Date	12 June 2017	